

## **BLANK PAGE**







Indian Standard

(Reaffirmed 1998)

## SPECIFICATION FOR RETAINERS, MATRIX, DENTAL, NO. 1 AND 2

- 1. Scope Dimensions and other requirements for matrix retainers No. 1 and 2 used in dentistry.
- 2. Shapes and Dimensions As shown in Fig. 1 and 2.
- 3. Material Stainless steel conforming to Designation 22Cr13 of Schedule V of IS: 1570-1961 'Schedules for wrought steels for general engineering purposes'.
- 4. Workmanship and Finish
- 4.1 The instrument shall be free from defects which detract from its appearance or impair its service-ability.
- 4.2 All surfaces and edges shall be smooth to prevent pinching or scratching of mouth tissue.
- 4.3 The instrument shall be suitably passivated and polished bright.
- 5. Heat Treatment The instrument shall be uniformly hardened and tempered to 370 to 410 HV.
- 6. Corrosion Resistance Test The instrument shall be tested as given in 8 of IS:3887-1966 'General requirements for cutting type dental instruments'.
- 7. Marking The instrument shall be marked with the number of the instrument; manufacturer's name, initials or trade-mark; the letters 'SS'; and the country of manufacture.
- 7.1 ISI Certification Marking Details available from the Indian Standards Institution, New Delhi 1.
- 8. Packing As agreed to between the purchaser and the supplier.





All dimensions in millimetres.

FIG. 2 RETAINER, MATRIX, DENTAL, NO. 2